

See discussions, stats, and author profiles for this publication at: <https://www.researchgate.net/publication/390286062>

## ORIGINALARTICLE

Article · March 2025

DOI: 10.57234/jiimc.march25.1785

CITATIONS

0

READS

38

4 authors, including:



**Akhtar Sohail Chughtai**

Chughtai Lab

62 PUBLICATIONS 194 CITATIONS

[SEE PROFILE](#)



**Omar Chughtai**

The University of Texas Health Science Center at Houston

42 PUBLICATIONS 115 CITATIONS

[SEE PROFILE](#)



**Hijab Batool**

Chughtai Lab

50 PUBLICATIONS 70 CITATIONS

[SEE PROFILE](#)

## ORIGINAL ARTICLE

## Clinical Evaluation of Serological Diagnostic Assays for COVID-19 Antibody Detection

Hijab Batool<sup>1</sup>, Muhammad Dilawar Khan<sup>2</sup>, Omar Rasheed Chughtai<sup>3</sup>, Akhtar Sohail Chughtai<sup>4</sup>, Shakeel Asharf<sup>5</sup>, Saqib Shahzad<sup>6</sup>

## ABSTRACT

**Objective:** This study was aimed to evaluate the laboratory performance of three different serological assays (Immunofluorescence Assay (IFA), Electrochemiluminescence Immunoassay (ECLIA) and Chemiluminescent Microparticle Immunoassay (CMIA) to see if they performed accurately according to the manufacturers' claims.

**Study Design:** Cross sectional study.

**Place and Duration of Study:** This study was conducted at Chughtai Institute of Pathology from 01<sup>st</sup> April to 30<sup>th</sup> May 2020.

**Materials and Methods:** Blood samples were collected from 75 adult male and female patients, 25 were pre pandemic samples and 50 were diagnosed cases of COVID-19 in whom sample was taken 21 days after they showed up symptoms. All cases were analyzed to detect the presence or absence of COVID-19 IgG antibodies using Immunofluorescence assay (IFA), electrochemiluminescence immunoassay (ECLIA) and Chemiluminescent Microparticle Immunoassay (CMIA). SPSS 23.0 and EP evaluator were used to assess sensitivity, specificity and Cohen's kappa.

**Results:** The study compares the effectiveness of three diagnostic methods (ECLIA, CMIA, and IFA) against PCR for detecting COVID-19 antibodies using Cohen's Kappa statistics. ECLIA showed the highest agreement with PCR (Kappa 0.748), followed by CMIA (Kappa 0.602), and IFA (Kappa 0.564), indicating that ECLIA is the most reliable method for detecting both positive and negative cases. The findings suggest variability in accuracy across these methods, with ECLIA being the most consistent.

**Conclusion:** Detection of anti-SARS-CoV-2 antibodies may act as a reliable diagnostic tool provided the assay is properly validated before use. Chemiluminescence immunoassay proves to be a better serological assay as compared to Electrochemiluminescence and Immunofluorescence assay.

**Key Words:** Chemiluminescence, Electrochemiluminescence, Immunofluorescence, SARS-Cov-2, Serological Assay.

## Introduction

In Wuhan City, Hubei Province, Central China, several patients with pneumonia of unknown etiology surfaced at the start of December 2019. The pneumonia known as coronavirus disease 2019 (COVID-19) has been shown through genome sequencing to be caused by a novel coronavirus (CoV) called severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2), formerly known as 2019 novel coronavirus (2019-nCoV).<sup>1</sup> The four genera of

Corona viruses are Alpha, Beta, Gamma, and Delta, and they are members of the "Corona Viridae" family.<sup>2,3</sup> Similar to SARS-CoV and MERS-CoV, the recently discovered SARS-CoV-2 virus is a member of the  $\beta$ -CoV B lineage.<sup>3</sup>

Once inside the epithelial cells, SARS-CoV-2 multiplies quickly and triggers a cytokine storm and immunological reaction that damages the pulmonary parenchyma. This hypercytokinemia results in multiple organ failure and acute respiratory distress syndrome due to its unchecked synthesis of proinflammatory cytokines.<sup>4</sup> Studies have revealed that cytokine storm syndrome occurred in severe cases of COVID-19, some of which deteriorated and died of multiple organ damage.<sup>5</sup>

SARS-CoV-2 outbreak and spread can only be controlled by rapid detection of the cases. Reverse transcriptase quantitative PCR (RT-qPCR) is a diagnostic tool based on nucleic acid sequencing

<sup>1,2,5</sup>Department of Chemical Pathology and Immunology/  
Histopathology<sup>3,4</sup>/Material Management<sup>6</sup>

Chughtai Institute of Pathology, Lahore, Pakistan

Correspondence:

Dr. Hijab Batool

Assistant Professor

Department of Chemical Pathology and Immunology

Chughtai Institute of Pathology, Lahore, Pakistan

E-mail: dr.hijabbatool@cli.edu.pk

Received: February 02, 2024 ; Revised: December 28, 2024

Accepted: March 05, 2025

with high sensitivity and specificity<sup>6</sup>. Apart from the diagnostic tests like RT-qPCR, serological tests have been in high discussion since the outbreak of this pandemic. A wide range of serological immunoassays with variable antibody specificities have been developed<sup>7,8</sup>. The COVID-19 virus has underscored the urgent need for reliable diagnostic antibody testing to accurately identify past infections and assess immunity levels within populations. Such testing is critical for guiding public health decisions, managing vaccination strategies, and understanding the virus's spread. Ensuring high accuracy in antibody tests helps prevent false results, thereby maintaining trust in public health measures. Accurate antibody testing is pivotal for monitoring and controlling the pandemic effectively. It is a major responsibility of clinical laboratories all around the world to validate these new methodologies before these techniques get introduced into routine clinical practice.<sup>9</sup>

In the given emergency of COVID-19, FDA issued relaxed regulatory guidelines to use SARS-CoV-2 serological assays to check the immune response of the population. Under these circumstances, it is now the duty of clinical laboratories to validate these assays rigorously to determine whether these assays perform accurately according to the package inserts. The serological assays being used around the world include rapid diagnostic tests (RDT), ELISA (Enzyme linked Immunosorbent assay), neutralization immunoassays and chemiluminescence. All these tests vary in the antigens they are designed to target e.g. Nucleocapsid protein (N Protein) or Spike protein (S Protein). This study aimed to detect the analytical performance of three different serological assays to detect COVID-19 IgG Antibody. The findings will help health care providers to identify a better serological assay for COVID-19 IgG antibody detection that is properly validated and solve many unanswered queries of the clinicians.

## Materials and Methods

This cross-sectional study was conducted at Chughati Institute of Pathology from 01<sup>st</sup> April to 30<sup>th</sup> May 2020. Ethical approval was obtained by the Institutional Review Board under letter number CIP/IRB/1029. Blood samples were collected from 75 adult male and female patients. Of these, 25 were pre-pandemic samples and 50 were COVID-19 cases

with confirmed diagnosis via RT-qPCR; the samples were obtained 21 days after the onset of symptoms. The pre-pandemic samples included specimens of healthy adult population collected in November 2019 and kept frozen at -80 ° Celsius. The healthy population was recruited according to WHO criteria (Constitution of the World Health Organization) after filling in the health questionnaire. These samples were collected as part of routine sample collection to establish a healthy sample pool for the laboratory's biobank. COVID-19 cases included in the study were admitted in high dependency corona units and hospital isolation wards and having mild to severe symptoms. COVID-19 patients with mechanical ventilation, asymptomatic cases, and those undergoing plasma infusion were not included in this research. In this study, we did not stratify the patients according to the severity of their symptoms. A volume of 3 ml blood was collected and centrifuged at 3000 RPM prior to analysis.

All samples were analyzed to detect the presence or absence of COVID-19 IgG antibodies using three different assays.

### 1. Immunofluorescence Assay (IFA) Lifotronic FA 160

This is a rapid diagnostic test detecting the presence of antibody on nitrocellulose membrane using immunofluorescence as test principle. Sample is added on the sample pad which moves forward through capillary force towards detection line of enveloped antigen (recombinant nucleocapsid protein) on the test strip. The strip is then placed into the incubation chamber of the reader for 15 minutes. If antibody is present, it will combine with antigen and fluorescence marker to form immune complexes along with a control line that can also produce fluorescence. After incubation, the test strip is placed into the test chamber and the result is displayed on the screen and printed as well. The whole process takes 20 minutes.

### 2. Elecsys Anti SARS-CoV-2 Assay- Roche Diagnostics

The electrochemiluminescence immunoassay (ECLIA) used in this study was for the qualitative detection of SARS-CoV-2-specific antibodies (IgG and IgM). The nucleocapsid (N) antigen recombinant protein is used in this assay to

measure the antibodies against SARS-CoV-2. The samples are classified as reactive (COI >1.00) or non-reactive (COI <1.00) based on a cut off index (COI) of 1.00.

### 3. SARS-CoV-2 IgG assay- Abbott Diagnostics

This is a chemiluminescent microparticle immunoassay (CMIA) for qualitative detection of IgG antibodies against SARS-CoV-2. This technique measures IgG antibodies against SARS-CoV-2 using nucleocapsid (N) antigen recombinant protein coated microparticles. The samples are classified as reactive (COI >1.40) or non-reactive (COI <1.40) based on a cut off index (COI) of 1.40.

All samples were analyzed on these assays after quality check and calibration. RT-PCR was used as a reference method. SPSS 23.0 was used to calculate percentages and frequencies, Cohen Kappa, sensitivity and specificity and Cohen Kappa. Cohen's Kappa is a statistical measure used to assess the agreement between two raters or diagnostic tests, beyond what would be expected by chance. When applied to sensitivity and specificity analysis, Cohen's Kappa helps evaluate the sensitivity, which measures the proportion of actual positives that are correctly identified by the tests and the specificity, which measures the proportion of actual negatives that are correctly identified. Cohen's Kappa evaluates if the agreement on cases between two tests exceeds what would be expected by chance. Cohen's Kappa values range from -1 to 1, where: Kappa ≤ 0: Indicates no agreement or agreement worse than chance, 0.01–0.20: Slight agreement, 0.21–0.40: Fair agreement, 0.41–0.60: Moderate agreement, 0.61–0.80: Substantial agreement and 0.81–1.00: Almost perfect or perfect agreement. Higher Kappa values suggest better consistency between the two methods, while values near or below zero indicate that the agreement might be due to chance.

## Results

The given results involved comparing three different diagnostic methods (ECLIA, CMIA, and IFA) against a reference standard (PCR) for detecting COVID-19 antibodies. The level of agreement between each method and PCR is assessed using Cohen's Kappa statistics, which helps determine how well these methods perform in identifying positive and

negative cases relative to the PCR results. Our study showed that ECLIA and PCR both identified 25 negatives and 42 positives and ECLIA incorrectly identified 8 positive cases as negative. There is substantial agreement between ECLIA and PCR results, with a Kappa value of 0.748. This indicates that ECLIA is a reliable method for COVID-19 detection compared to PCR, with high consistency in detecting both positive and negative cases. CMIA and PCR both identified 25 negatives and 36 positives and CMIA incorrectly identified 14 positive cases as negative. There is moderate agreement between CMIA and PCR results, as indicated by a Kappa value of 0.602. While CMIA shows a good level of agreement, it is less consistent than ECLIA, especially in identifying positive cases. There is a higher rate of false negatives compared to ECLIA. IFA and PCR both identified 25 negatives and 33 positives. IFA incorrectly identified 17 positive cases as negative. The Kappa value of 0.564 indicates a moderate level of agreement between IFA and PCR results. IFA has the lowest agreement among the three methods, with a higher tendency to miss positive cases (false negatives). This suggests that IFA is less reliable than ECLIA and CMIA in accurately detecting COVID-19 cases when compared to PCR. These findings suggest that ECLIA is the most effective among the three methods for diagnosing COVID-19, based on the agreement with PCR results. In this study, all three assays showed different negative and positive agreement along with Cohen's Kappa for the pre pandemic and post pandemic samples. Comparison of performance of all three assays is given in table I

**Table I: Sensitivity And Specificity of Assays Used for Serological Testing (n=75)**

Assay	Negative agreement (Specificity)	Positive agreement (Sensitivity)	Cohen's Kappa
IFA	59%	66%	0.564
ECLIA	100%	84%	0.748
CMIA	100%	72.0%	0.602

## Discussion

In this work, we compared and clinically assessed three commercially available tests for the detection of SARS-CoV-2 antibodies. Among all the tests used for the qualitative assessment of Anti-SARS-CoV-2 antibodies, the ECLIA test from Roche Diagnostics had the best specificity, followed by the Abbott

assay. IFA by Lifotronic demonstrated the lowest sensitivity and specificity and was linked to most false positive and false negative outcomes. The requirement for specialized equipment and skilled personnel to interpret results may also limit accessibility and consistency of results. Additionally, cross-reactivity with antibodies from other coronaviruses can cause false-positive results, while low antibody titers in early or mild infections may lead to false negatives.

Given these limitations, some labs opted to use Electrochemiluminescence Immunoassay (ECLIA) and Chemiluminescence Microparticle Immunoassay (CMIA) for COVID-19 antibody detection instead of Immunofluorescence Assay (IFA). ECLIA and CMIA offer higher throughput, automation capabilities, and improved sensitivity and specificity, leading to more consistent and faster results. These methods also minimize human error associated with manual interpretation in IFA and are less labor-intensive, making them ideal for large-scale testing in clinical settings.

A recent American study found that the Abbott assay using CMIA and ECLIA was more reliable than rapid tests after 14 days of symptom onset. Specifically, the study showed that the Abbott SARS-CoV-2 assay had a diagnostic sensitivity of 93.8% and a specificity of 99.4% after this period.<sup>10</sup> The variability in study results compared to manufacturers' claims can be attributed to several factors. Patients often present overlapping clinical scenarios and are under different treatment plans. Moreover, testing on hospitalized patients with immunodeficiencies and comorbidities can affect results. Some manufacturers also measure assay sensitivity from the time of RT-qPCR positivity rather than from the onset of symptoms, potentially leading to an overestimation of sensitivity.<sup>10</sup> In our study, serological assessments in COVID-19 patients were conducted 21 days post-symptom onset, a point at which the maximum serological response is typically observed.<sup>11</sup>

None of the pre-pandemic samples in our study showed seropositivity when analyzed using CMIA and ECLIA, whereas IFA showed considerable false positives, likely due to the assay's low specificity. Some studies have found false positive serological results in pre-pandemic confirmed cases of seasonal

coronavirus, attributing this to the structural homology between seasonal coronaviruses and SARS-CoV-2, which can affect the specificity of assays like ELISA. In regions with low prevalence, such as Pakistan, where the current SARS-CoV-2 attack rate is 2.3 per 100,000 population, the necessity of high specificity becomes crucial to achieve a high positive predictive value. This is important for public health strategies and testing protocols because a low specificity can lead to an overestimation of infection rates, thereby affecting resource allocation and intervention strategies. Understanding the regional epidemiology helps inform the selection of assays with optimal specificity and sensitivity to ensure accurate diagnosis and effective public health response.<sup>13</sup> In this case, obtaining a high positive predictive value necessitates a serological test with good specificity.<sup>14</sup> According to the FDA, the performance of an assay depends on the population prevalence, and in low-prevalence populations, a single antibody test may not be sufficient to differentiate true positives from false positives. In clinical settings, especially among hospitalized patients with comorbidities, the immune response can be atypical, potentially leading to altered assay performance. These patients may have impaired immune responses due to their underlying conditions or treatments, which can affect the production of antibodies, thereby influencing both sensitivity and specificity of the tests. As a result, antibody tests may produce higher rates of false positives or false negatives in these environments, highlighting the need for confirmatory testing or using assays with high specificity to ensure accurate diagnosis. Understanding the impact of these clinical variables is crucial for interpreting test results and making informed public health decisions.<sup>15</sup>

Horber et al., assessed several SARS-CoV-2 serological tests for antibody detection and discovered that the assays had a good level of sensitivity and specificity at least 14 days after PCR positive. The investigators found the diagnostic sensitivity of Siemens to be highest compared to Roche and Euroimmun.<sup>16</sup> Clinical evaluation of different serological assays for SARS-CoV-2 reveals that chemiluminescence immunoassay (CLIA) show 100% specificity (sample collected 12 days post symptom onset), RDT show 90.3% clinical specificity



(sample collected 33 days post symptom onset) while the diagnostic sensitivity while using ELISA as assay of choice was 66.7% in the early phase of disease.<sup>11,17,18</sup> A comparison between Roche ECLIA and DiaSorin Liaison CMIA SARS-CoV-2 IgG assay is in accordance with our study demonstrated specificities of 100% and 98.9% with an overall agreement of 99% with RT-PCR.<sup>19</sup> Another recently published study compared the diagnostic sensitivities of Spike protein based serological assay and Nucleocapsid protein-based assay and found that later having high sensitivity (77.8%).<sup>20</sup> The assays used in our study were also Nucleocapsid protein based. An American study comparing the specificity of SARS-CoV-2 IgA and IgG assay (Euroimmun) found IgG assay to have higher specificity (97%) as compared to IgA assay (81%).<sup>21</sup> Our study is also in concordance with a recent evaluation of three commercial SARS-CoV-2 serological assays i.e. Abbott IgG, Roche total antibody and DiaSorin IgG. In this study, the nucleocapsid antibody test (Abbott and Roche) showed higher sensitivity as compared to Spike protein Antibody test (DiaSorin).<sup>22</sup> The Infectious Diseases Society of America (IDSA) advises using serology to diagnosis patients who have a high suspicion of COVID-19 but who test negative for the virus by RT-PCR.<sup>23</sup> Using a variety of immunochromatographic tests, Demey et al. showed that antibodies against SARS-CoV-2 may be found approximately 10 days after the onset of symptoms where as Thevarajan recorded an increase in the antibody production from 7 to 20 days after beginning of the disease.<sup>24,25</sup> Many studies suggest that serological assays can act as reliable diagnostic tool for identification of SARS-CoV-2 infection as well as help to determine the immune status of the population since they have high sensitivity and specificity. Previously it was debated that serological assays cannot be used to diagnose SARS-CoV-2 infection, however, recently many studies postulate that antibody assays can be used to diagnose COVID-19.<sup>26</sup> Serological assays having high sensitivity and specificity can be used to screen asymptomatic cases, early diagnosis of the infection (with the help of IgM only), and monitoring response during treatment<sup>27-29</sup>. However, negative tests cannot be used to exclude the infection considering the fact that patients may have been recently

exposed to the virus<sup>30</sup>. Another fact to be considered is assay showing cross reactivity to non-SARS-CoV-2 proteins. Risk assessment for healthcare personnel, epidemiological surveys, and vaccine research can all benefit from the use of high sensitivity serological testing. However, appropriate validation of the assays' diagnostic accuracy is necessary for all these surveys and investigations.

## **Conclusion**

Besides RT-PCR, to confirm the presence of COVID-19 in the suspected cases, detection of anti-SARS-CoV-2 antibodies may act as a reliable diagnostic tool provided the assay is properly validated before use. Electrochemiluminescence immunoassay proves to be a better serological assay as compared to chemiluminescence and Immunofluorescence assay. Assays with higher diagnostic sensitivity and specificity can overcome the RT-PCR limitations helping to diagnose asymptomatic carriers and false negative RT-PCR cases.

## **Limitation**

- coronavirus genomic 3' and 5' ends. *Virus Res.* 2015;206:120–133. doi: 10.1016/j.virusres.2015.02.025.
4. Villar J, Zhang H, Slutsky AS. Lung repair and regeneration in ARDS: role of PECAM1 and Wnt signaling. *Chest.* 2019;155:587–594. doi: 10.1016/j.chest.2018.10.022.
  5. Chen N, Zhou M, Dong X, Qu J, Gong F, Han Y. Epidemiological and clinical characteristics of 99 cases of 2019 novel coronavirus pneumonia in Wuhan, China: a descriptive study. *Lancet.* 2020;395:507–513. doi: 10.1016/S0140-6736(20)30211-7.
  6. Corman VM, Landt O, Kaiser M, Molenkamp R, Meijer A, Chu DKW. Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. *Euro Surveill.* 2020;25 doi: 10.2807/1560-7917.ES.2020.25.3.2000045.
  7. Vashist SK. In Vitro Diagnostic Assays for COVID-19: Recent Advances and Emerging Trends. *Diagnostics (Basel)* 2020;10:202.
  8. Winter AK, Hegde ST. The important role of serology for COVID-19 control. [epub ahead of print] *Lancet Infect Dis* 2020 Apr 21 as doi: 10.1016/S1473-3099(20)30322-4.
  9. Kirkcaldy RD, King BA, Brooks JT. COVID-19 and Postinfection Immunity: Limited Evidence, Many Remaining Questions. [epub ahead of print] *JAMA* 2020 May 11 doi: 10.1001/jama.2020.7869.
  10. Tang M, Hock K, Logsdon N, Hayes J, Gronowski A, Anderson N, et al. Clinical performance of two SARS-CoV-2 serologic assays. *Clin Chem.* 2020;66(8):1055–62.
  11. Zhao J, Yuan Q, Wang H, et al. Antibody responses to SARS-CoV-2 in patients of novel coronavirus disease 2019. *Clin Infect Dis* 2020. doi: 10.1101/2020.03.02.20030189.
  12. To KK-W, Tse OT-Y, Leung W-S, Tam AR, Wu T-C, Lung DC, et al. Temporal profiles of viral load in posterior oropharyngeal saliva samples and serum antibody responses during infection by SARS-CoV-2: an observational cohort study. *Lancet Infect Dis.* 2020;20(5):565–74.
  13. Abid K, Bari YA, Younas M, Tahir Javaid S, Imran A. Progress of COVID-19 epidemic in Pakistan. *Asia Pac J Public Health.* 2020;32(4):154–6. doi: 10.1177/1010539520927259.
  14. Farnsworth CW, Anderson NW. SARS-CoV-2 serology: much hype, little data. *Clin Chem.* 2020;66(7):875–7. doi: 10.1093/clinchem/hvaa107.
  15. EUA Authorized Serology Test Performance [Internet]. U.S. Food and Drug Administration. 2020 [cited 13 October 2020]. Available from: <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/eua-authorized-serology-test-performance>.
  16. Hörber S, Soldo J, Relker L, Jürgens S, Guthier J, Peter S et al. Evaluation of three fully automated SARS-CoV-2 antibody assays. *Clinical Chemistry and Laboratory Medicine (CCLM).* 2020;0(0).
  17. Infantino M, Grossi V, Lari B, Bambi R, Perri A, Manneschi M, et al. Diagnostic accuracy of an automated chemiluminescent immunoassay for anti-SARS-CoV-2 IgM and IgG antibodies: an Italian experience. *J Med Virol.* 2020.
  18. Li, Z., Yi, Y., Luo, X., Xiong, N., Liu, Y., Li, S., et al. Development and clinical application of a rapid IgM-IgG combined antibody test for SARS-CoV-2 infection diagnosis. *J. Med. Virol.* 2020.
  19. Merrill AE, Jackson JB, Ehlers A, Voss D, Krasowski MD. Head-to-Head Comparison of Two SARS-CoV-2 Serology Assays [published online ahead of print, 2020 Jul 27]. *J Appl Lab Med.* 2020;jfaa125. doi:10.1093/jalm/jfaa125.
  20. Kohmer N, Westhaus S, Rühl C, Ciesek S, Rabenau H. Brief clinical evaluation of six high-throughput SARS-CoV-2 IgG antibody assays. *Journal of Clinical Virology.* 2020;129:104480.
  21. Wheeler SE, Shurin GV, Keetch C, et al. Evaluation of SARS-CoV-2 prototype serologic test in hospitalized patients [published online ahead of print, 2020 Aug 26]. *Clin Biochem.* 2020;S0009-9120(20)30816-X. doi:10.1016/j.clinbiochem.2020.08.008.
  22. Sarah E. Turbett, Melis Anahtar, Anand S. Dighe, Wilfredo Garcia Beltran, et al. Branda Journal of Clinical Microbiology Oct 2020, JCM.01892-20; DOI: 10.1128/JCM.01892-20.
  23. Idsociety.org. 2020 [cited 13 October 2020]. Available from: <https://www.idsociety.org/globalassets/idsa/public-health/covid-19/idsa-covid-19-antibody-testing-primer.pdf>.
  24. Demey B, Daher N, François C, Lanoix JP, Duverlie G, Castelain S, et al. Dynamic profile for the detection of anti-SARS-CoV-2 antibodies using four immuno-chromatographic assays. *J Infect.* 2020 Aug;81(2):e6–e10. doi: 10.1016/j.jinf.2020.04.033.
  25. Thevarajan I, Nguyen THO, Koutsakos M, Druce J, Caly L, van de Sandt CE, Jia X, Nicholson S, Catton M, Cowie B, Tong SYC, Lewin SR, Kedzierska K. Breadth of concomitant immune responses prior to patient recovery: a case report of non-severe COVID-19. *Nat Med.* 2020 Apr;26(4):453–455. doi: 10.1038/s41591-020-0819-2.
  26. Xiang F, Wang X, He X, Peng Z, Yang B, Zhang J, Zhou Q, Ye H, Ma Y, Li H, Wei X, Cai P, Ma WL. Antibody Detection and Dynamic Characteristics in Patients with Coronavirus Disease 2019. *Clin Infect Dis.* 2020 Nov 5;71(8):1930–1934. doi: 10.1093/cid/ciaa461.
  27. Li Z, Yi Y, Luo X, Xiong N, Liu Y, Li S, et al. Development and clinical application of a rapid IgM-IgG combined antibody test for SARS-CoV-2 infection diagnosis. *J Med Virol.* 2020 Sep;92(9):1518–1524. doi: 10.1002/jmv.25727.
  28. Wu X, Fu B, Chen L, Feng Y. Serological tests facilitate identification of asymptomatic SARS-CoV-2 infection in Wuhan, China. *J Med Virol.* 2020 Oct;92(10):1795–1796. doi: 10.1002/jmv.25904.
  29. Pan Y, Li X, Yang G, Fan J, Tang Y, Zhao J, et al. Serological immunochromatographic approach in diagnosis with SARS-CoV-2 infected COVID-19 patients. *J Infect.* 2020 Jul;81(1):e28–e32. doi: 10.1016/j.jinf.2020.03.051.
  30. Cheng MP, Papenburg J, Desjardins M, Kanjilal S, Quach C, Libman M, et al. Diagnostic Testing for Severe Acute Respiratory Syndrome-Related Coronavirus 2: A Narrative Review. *Ann Intern Med.* 2020 Jun 2;172(11):726–734. doi: 10.7326/M20-1301.

**CONFLICT OF INTEREST**

Authors declared no conflicts of Interest.

**GRANT SUPPORT AND FINANCIAL DISCLOSURE**

Authors have declared no specific grant for this research from any funding agency in public, commercial or nonprofit sector.

**DATA SHARING STATEMENT**

The data that support the findings of this study are available from the corresponding author upon request.

This is an Open Access article distributed under the terms of the Creative Commons Attribution- Non-Commercial 2.0 Generic License.

.....